

Setting up the Talos Arctica electron microscope and Gatan K2 direct detector for high-resolution cryogenic single-particle data acquisition

Gabriel C. Lander (✉ glander@scripps.edu)

Lander Lab, The Scripps Research Institute

Mark A. Herzik, Jr.

Mengyu Wu

Method Article

Keywords: cryo-EM, electron microscopy, 200 keV, Talos Arctica, Gatan K2

Posted Date: October 12th, 2017

DOI: <https://doi.org/10.1038/protex.2017.108>

License: © ⓘ This work is licensed under a Creative Commons Attribution 4.0 International License.

[Read Full License](#)

Abstract

It has been determined that a Talos Arctica transmission electron microscope (TEM) is capable of resolving frozen-hydrated macromolecules of various sizes and internal symmetries to better than 3 Å resolution using single particle methodologies. This protocol describes how to maximize parallel illumination conditions in a Talos Arctica TEM operating at 200 keV for high-resolution cryogenic data acquisition for structure determination.

Introduction

When collecting data using a two-condenser lens TEM, it is imperative that the lenses are set to illuminate the sample with a maximally parallel beam in order to prevent changes in defocus from causing magnification changes. Larger defocus changes produce a greater degree of magnification distortion. Although the magnification changes resulting from non-parallel illumination are subtle, they will nonetheless degrade the high-resolution information of a reconstruction when thousands of particles from different defoci are averaged together. This protocol assumes that: 1) Gun alignments have already been performed on the microscope (note that only service engineers should change the gun settings on an X-FEG), 2) the Microscope User Interface and Digital Micrograph software are installed on the scope and camera computers, respectively, 3) grids amenable to data collection have already been prepared and loaded into the microscope, and 4) SA-mode magnifications use NanoProbe mode, all M and LM magnifications use MicroProbe.

Reagents

Holey EM grids containing frozen-hydrated sample

Equipment

Talos Arctica TEM operating at 200 keV equipped with a Gatan K2 Summit direct electron detector and 4K × 4K CETA 16M CCD camera (or similar), operating with the following microscope settings: - Extraction voltage: 4150 V - Gun lens setting: 8 - Spot size 3 (or 4 depending on desired exposure rate, see B.8) - C1 aperture: 2000 μm - C2 aperture: 70 μm - Objective aperture: 100 μm 500 nm crossed lines grating replica calibration grid (Ted Pella, Inc. Product # 677 or similar). Autogrid C-clip rings (FEI, Product #1036173) and Autogrid C-clips (FEI Product #1036171) for Talos Arctica TEM autoloader.

Procedure

****A. K2 gain corrections**** 1) Navigate to an empty square and change to a magnification within the SA lens series (e.g. SA 45000x). 2) In Digital Micrograph, go to Camera → Prepare Gain Reference. Follow prompts. 3) Collect new hardware dark references every day of data collection ****B. Parallel illumination**** 1) Move to an intact square on the calibration grid and determine eucentric height (Z-shift). 2) With the

objective aperture removed determine proper focus at a magnification of SA 45000x using Digital Micrograph to display the Fourier Transform of a live image. 3) Center the C2 aperture (70 μm) and stigmatize the C2 lens. 4) At proper focus and eucentric height, align the beam tilt pivot points. 5) At a magnification of 120000x determine the rotation center alignment of the objective lens. 6) Change to SA 45000x magnification. While in diffraction mode, insert the objective aperture (100 μm) and move the aperture off-center so that the edge of the aperture is overlapping the gold powder diffraction rings. 7) Adjust using the focus knob until the edge of the objective aperture is as sharp/crisp as possible. Then, using the intensity knob, minimize the width of the gold powder diffraction rings. 8) Make note of the beam intensity value, as this value is to be used for exposure image acquisition under conditions of parallel illumination. NOTE: Exposure rate adjustments must be performed using changes in spot size rather than changes in beam intensity. If changes in spot size, exposure magnification, C2 or objectives aperture are made, then steps B.1-7 need to be repeated. 9) Center the objective aperture and stigmatize the objective lens.

****C. Determine exposure rate****

- 1) Navigate to an empty square and set scope to 45000x magnification and beam intensity value as determined in step B.8
- 2) In Digital Micrograph, click on the settings icon in the "Camera View" window, and make sure that the binning = 1, and click on "Full CCD".
- 3) Set an exposure time of 1 second and click on "start view". The exposure rate (in $\text{e}^-/\text{pixel}/\text{s}$) will show up in the Camera Monitor window. A value of 2-5 $\text{e}^-/\text{pixel}/\text{s}$ is desired. NOTE: Exposure rate adjustments must be performed using changes in spot size rather than changes in beam intensity. If changes in spot size, exposure magnification, C2 or objectives apertures are made, then steps B.1-8 need to be repeated.
- 4) Using the determined exposure rate, adjust the total movie exposure time to yield a cumulative exposure of $\sim 50\text{-}65 \text{ e}^-/\text{\AA}^2$. Adjust the per-frame exposure time (e.g. 100-250 ms) to yield a per-frame cumulative exposure greater than $0.8 \text{ e}^-/\text{\AA}^2/\text{frame}$, while maximizing the number of frames for better correction of per-frame movement.

****D. Coma-free alignment****

Coma-free alignments were performed as previously described (Glaeser et al. 2011), and steps below carried out using the Legikon software. This procedure is done to confirm that rotation center alignment done in step B.5 was performed correctly.

- 1) Move to an intact square on the calibration grid and determine eucentric height and proper focus.
- 2) Create a preset that uses the same magnification and intensity settings as the exposure preset. Set this preset to use a defocus of $-3\text{e}^{-7} \mu\text{m}$ and acquire 924×924 (binned by 2) images with a 2 second exposure.
- 3) Go to the "Beam Tilt Image" node and set these parameters in the settings:
 - Tableau Type: beam tilt series-power
 - Beam Tilt: 0.005 radians
 - Number of tilt directions: 4
 - Start angle: 0
 - Correlation type: phase
 - Tableau binning: 2
 - beam tilt count: 1
- 4) Click on the "Simulate target" icon in the Beam Tilt Image node.
- 5) The opposing images in the resulting tableau should have approximately the same amount of astigmatism and Thon ring spacing. If adjustments are necessary, we recommend repeating the rotation center alignment (step B.5).

References

Glaeser RM, Typke D, Tiemeijer PC, Pulokas J, & Cheng A (2011) Precise beam-tilt alignment and collimation are required to minimize the phase error associated with coma in high-resolution cryo-EM. J

Struct Biol 174(1):1-10. Suloway C, et al. (2005) Automated molecular microscopy: the new Legimon system. J Struct Biol 151(1):41-60.