

Improved access in minimally invasive temporomandibular joint surgery through a novel endaural template.

Matthias Krause

Leipzig University: Universität Leipzig

Mohammad Kamal

Kuwait University

Daniel Kruber

HTWK Leipzig: Hochschule für Technik Wirtschaft und Kultur Leipzig

Dirk Halama

Leipzig University: Universität Leipzig

Thomas Hierl

HELIOS Vogtland-Klinikum Plauen

Bernd Lethaus

Leipzig University: Universität Leipzig

Alexander K. Bartella (✉ alexander.bartella@medizin.uni-leipzig.de)

Leipzig University: Universität Leipzig <https://orcid.org/0000-0002-3545-1341>

Technical advance

Keywords: TMJ surgery, template, endaural template, minimally invasive temporomandibular joint surgery

Posted Date: January 26th, 2021

DOI: <https://doi.org/10.21203/rs.3.rs-101466/v2>

License: © ⓘ This work is licensed under a Creative Commons Attribution 4.0 International License.

[Read Full License](#)

Version of Record: A version of this preprint was published on February 19th, 2021. See the published version at <https://doi.org/10.1186/s12893-021-01098-2>.

Abstract

Background

Digitally designed surgical templates for minimally invasive temporomandibular joint (TMJ) surgery (MITMJS) are a promising tool for improving the safety of these procedures. Given the TMJ anatomy, the template fitting and intraoperative overview are the most important issues for a safe surgery. This article is a technical advance article that aims to describe an endaural surgical template based on the Moses approach as a possible solution in TMJ surgery.

Methods:

Three patients with internal derangement were treated with the guidance of a MITMJS template based on cone beam computed tomography (CBCT) and a surface imprint of the periauricular region. None of the patients needed an additional open surgical procedure. Fitting of the templates was judged in terms of position and rotational stability. Surgical side effects and complications were recorded for each patient.

Results:

The template design and clinical use were satisfactory for MITMJS. The templates showed satisfying fit and good visibility. In the study cohort, no bleeding, facial nerve injury, or other complications occurred after the procedure, and no visible scars were noted postoperatively.

Conclusion:

Our feasibility report on template-guided MITMJS shows a promising new application of templates. It points to improved access in arthroscopy or arthrocentesis of TMJ surgery through endaural access with an increased level of safety during surgery.

Background

The advocacy of minimally invasive approaches to temporomandibular joint (TMJ) surgery was first published in 1975 [1]. The proposed advantages of such controlled access include lower surgical morbidity, faster surgical recovery, and a decreased chance of complications, such as facial nerve injury, bleeding, perforation of the external acoustic meatus or the articular cavity, as described in previous studies. [1-3] However, proper conduction of minimally invasive treatment requires an advanced level of expertise in the field of minimally invasive temporomandibular joint surgery (MITMJS). To simplify and standardize the approach, we described the use of a computer-aided designed and computer-aided manufactured (CAD-CAM) template for TMJ surgery in 2019, yet several modifications have been proposed to this promising technology. [4] Most of the described modifications require 3D imaging (e.g., CT, cone beam CT, MRI) of the corresponding TMJ and the surrounding tissue, in addition to an optical three-dimensional (3D) scan of the face. Out of these data, a stereolithic template was designed. Given

the TMJ anatomy, the fitting and unique placement of the template is of utmost importance to safely conduct the surgery.

To ensure proper fitting of the 3D surgical template, several designs have been proposed. Some surgeons prefer extensions of the template to the zygomatic bone and forehead, while others rely on occlusal and dental stabilization methods. Each design has its own drawbacks, as the design may position the fixation out of the primary field of view of the surgeon, so if the template dislocates slightly, this might go unnoticed, leading to errors. In addition, dental fixation of the template may lead to increased exposure to bacteria of the oral cavity. Another drawback in the currently described templates is that they are designed to be utilized only through the preauricular approach. This risks accidentally perforating the external acoustic meatus. Additionally, visualization and instrumentation of the lateral and medial joint landmarks might be difficult to handle due to anatomical limitations. This is a known surgical problem in arthroscopic TMJ surgery, which was addressed by Moses and Poker in 1989 [5]. They described an enaural approach to the TMJ leading to an increased field view and allowing for improved surgical handling.

In our study, we evaluated the conductivity of TMJ surgery using a CAD/CAM-based template through Moses' endural approach.

Materials

Patients and selection

This study was designed as a technical advance article. It was approved by the Local Ethics Committee at Leipzig University Hospital following the Declaration of Helsinki on medical protocols and ethics (Eth-30/17, 12/06/2017). Three patients of a tertiary care center who needed MITMS were included in the study from March to August 2020. The patients were diagnosed with an internal derangement (ID) of the TMJ, which was classified according to the Wilkes classification system [6]. All patients underwent TMJ surgery using the guided endaural template. None of the patients needed additional open surgical procedures. Follow-up time was three months. Outcome variables were categorical. We noted fitting of the templates, which was judged by the surgeon in terms of position and rotational stability. Furthermore, surgical side effects and complications were recorded for each patient. Patients with effect modifiers or a potential confounder, which would require an additional open approach, were excluded from the study because the template would no longer fit.

Surgical procedure

The surgical procedure was conducted under sedation or general anesthesia. The template was positioned in the ear, and the skin was marked through the pilot channel for the desired incision after

insertion of the trocar. The template was temporarily removed, and the subcutaneous tissue was spread with fine-point scissors. After the guide had been repositioned, while the mandible was distracted downward and forward, the anterior wall of the external auditory canal was perforated with the sharp trocar up to the capsule (endaural access). Next, the endoscope was inserted through the other pilot channel after marking and incising the skin (Figure 6). The authors preferred a 0° arthroscopic cannula for the procedures.

Template manufacturing

The production of an endaural template was conducted according to the following four steps:

1. A 3D data set of the patient's skull was created. This was done by computed tomography (CT), cone beam computed tomography (CBCT), or magnetic resonance imaging (MRI). In our cohort, CBCT data sets were used for all patients (Kodak 9500 3D; Carestream Health, Toulouse, France).
2. The surface imprint of the end- and preauricular region was assessed. A precise acquisition was necessary for the template's position and rotational stability (Figure 1). The external acoustic meatus was used as a "key lock structure", and an impression had to be taken using silicone material (Omnisil; Omnident Dental, Rodgau, Germany). The transfer of the imprint to a digital model was performed with CBCT from the plaster model (Figure 2) due to the availability of the device and ease of image fusion (Figure 3).
3. The data sets were fused in planning software (Facial Analysis Tool (FAT)). It was then possible to create the template with two pilot channels for the endoscope and the manipulation instruments from the superior-posterolateral and endaural approaches. The alignment was directed to the upper joint space (Figure 4).
4. Finally, the template was manufactured using a 3D printer (Formlab 2; Formlabs Inc., Somerville, USA) and a CE-certified biocompatible photopolymer resin (Dental SG Resin; Formlabs, Somerville, USA) (figure 5). Due to the small number of patients, no statistical analysis was performed.

Results

The data for all patients was complete. Three patients with internal derangement of three joints were treated with template guidance. The mean age was 51 years. Treatments and procedures are listed in Table 1. One patient was treated with lavage followed by injections of platelet-rich fibrin (PRF). One patient suffered from Wilkes class II, and two suffered from Wilkes class IV (Table 1). No bleeding, facial nerve injury, or other complications occurred after the procedure. The fitting of the template was

satisfactory. It was stable in terms of positioning and rotation. The template workflow took an additional 120 minutes.

Table 1: Patient data and performed surgical procedures (n = 3). The number of patients, sex, age, diagnosis, and procedure are shown.

Gender	Age	Diagnosis	Wilkes stage	Side treated	Procedure
Female	34	Internal derangement	II	left	arthroscopy, lavage
Female	67	internal derangement	IV	right	arthroscopy, lavage
Male	52	internal derangement	IV	left	arthroscopy, lavage, PRF-injection
Average	51				

Discussion

Minimally invasive surgery is widely utilized and sometimes regarded as the gold standard of treatment in several fields of medicine [7-9]. However, it is difficult in TMJ surgery due to the close proximity of anatomical structures. Several approaches are described in the literature for MITMJ surgery. The major drawbacks of the common anterior approaches are potential complications such as facial nerve injury or otic injury [10]. A solution to this problem was suggested by Moses et al. and their endaural approach to the TMJ [5].

Until now, this approach could not be transferred to a template. The major difficulties for transferring the pilot channels appeared to be the insufficient detection of the internal acoustic meatus by the face scanner, which is used in planning, a problem that we also faced in the treatment process. To solve this problem, we used a silicone impression of the external meatus. It has the drawback that the workflow is not entirely digital. However, the impression material can be easily digitalized, which leads to a perfect fit of the template. Until further improvement of 3D scanning is achieved, it remains the only reliable option to capture the anatomical morphological details.

Another drawback of the templates that were described in the literature is the potential for intraoperative dislocation and loss of positional accuracy. Most templates are designed with wings that reach the zygomatic bone or the forehead. However, given the elasticity of the skin and subdermal tissue, intraoperative dislocation might occur. Due to the design of the templates, dislocation might appear without the surgeon being aware of it. Because of the anatomical proximity to neighboring structures, the dislocation potentially leads to accidental nerve or vessel injury or perforations of the temporomandibular fossa or the acoustic meatus. The use of the external acoustic meatus as the locking structure for the template allows reliable and solid fixation. Furthermore, the surgeon has the complete template in their field of vision to detect a potential dislocation.

From the surgical point of view, Moses and Poker's technique offers a safe and considerably easy approach for treating the mandibular joint. The described advantages, such as facial nerve protection, can be confirmed in template-guided minimally invasive approaches without visible scars.

The bias and limitations of the study are attributed mostly to the small study sample. Furthermore, we believe that a fully digital workflow with a 3D scan of the endaural region might be more favorable to silicone impressions due to the less time needed.

Conclusion

Our feasibility report on template-guided MITMJS shows its usefulness and ability to handle new applications of templates. It proposes improved and safe access in arthroscopy or arthrocentesis of temporo-mandibular joint surgery through a novel endaural template.

Abbreviations

MITMJS - minimally invasive temporomandibular joint surgery

TMJ - temporomandibular joint

CBCT - cone beam computed tomography

CAD-CAM - computer aided manufactured

3D - three-dimensional

ID internal derangement

MRI - magnetic resonance imaging

FAT - Facial Analysis Tool

PRF - platelet rich fibrin

Declarations

Ethics approval and consent to participate

This study was approved by the Ethics Committee of the Berlin Medical Chamber (Eth-30/17, 12/06/2017). This report followed the Declaration of Helsinki on medical protocols and ethics. Written informed consent to participate was obtained from all patients.

Consent for publication

All authors consent for publication. Written informed consent to publish personal data and use their images was obtained by all patients.

Availability of data and material

Data can be provided. All data can be shared.

Competing interests

The authors declare that they have no conflicts of interest.

Funding

The author did not obtain funding for the conduction and/or presentation of the study.

Authors contributions

All authors have read and approved the manuscript. All contributed significantly to the work. M. K., D. G., D. H. and T. H. designed the template. M. K., B. L. and A. B. had a major role in study organization and draft writing.

Acknowledgments

Not applicable.

References

1. Onishi M: **[Arthroscopy of the temporomandibular joint (author's transl)]**. *Kokubyo Gakkai Zasshi* 1975, **42**(2):207-213.
2. Gonzalez-Garcia R, Rodriguez-Campo FJ, Escorial-Hernandez V, Munoz-Guerra MF, Sastre-Perez J, Naval-Gias L, Gil-Diez Usandizaga JL: **Complications of temporomandibular joint arthroscopy: a retrospective analytic study of 670 arthroscopic procedures**. *J Oral Maxillofac Surg* 2006, **64**(11):1587-1591.
3. Gonzalez-Garcia R: **The current role and the future of minimally invasive temporomandibular joint surgery**. *Oral Maxillofac Surg Clin North Am* 2015, **27**(1):69-84.
4. Krause M, Dorfler HM, Kruber D, Humpfner-Hierl H, Hierl T: **Template-based temporomandibular joint puncturing and access in minimally invasive TMJ surgery (MITMJS) - a technical note and first clinical results**. *Head Face Med* 2019, **15**(1):10.
5. Moses JJ, Poker ID: **Temporomandibular joint arthroscopy: the endaural approach**. *Int J Oral Maxillofac Surg* 1989, **18**(6):347-351.

6. Wilkes CH: **Internal derangements of the temporomandibular joint. Pathological variations.** *Arch Otolaryngol Head Neck Surg* 1989, **115**(4):469-477.
7. Lin HF, Chen YD, Chen SC: **Value of diagnostic and therapeutic laparoscopy for patients with blunt abdominal trauma: A 10-year medical center experience.** *PLoS One* 2018, **13**(2):e0193379.
8. Lu Q, Qu H, Liu C, Wang S, Zhang Z, Zhang Z: **Comparison of Laparoscopy and Laparotomy in Surgical Staging of Apparent Early Ovarian Cancer: 13-year Experience.** *Medicine (Baltimore)* 2016, **95**(20):e3655.
9. Al-Juburi A, Granger S, Barnes J, Voeller G, Beech D, Amiri H, Abell TL: **Laparoscopy shortens length of stay in patients with gastric electrical stimulators.** *JSLs* 2005, **9**(3):305-310.
10. Soni A: **Arthrocentesis of Temporomandibular Joint- Bridging the Gap Between Non-Surgical and Surgical Treatment.** *Ann Maxillofac Surg* 2019, **9**(1):158-167.

Figures



Figure 1

The conventional imprint with polyvinylsiloxane. Detailed capture of the three-dimensional external meatus is possible.




Figure 2

Converting the impression to a virtual three-dimensional data set that can be imported to planning software. In this case, a CBCT-scan was used.



Figure 3

Import of the previously generated 3D data set and design of the template. It was planned in FAT software. The view with skin surface is displayed.




Figure 4

Planning the pilot channels in FAT software. View in bone surface and direction of the pilot channels. They end at the TMJ capsule.



Figure 5

The final template 3D printed. They are composed of biocompatible photopolymer resin.



Figure 6

Intraoperative situation after insertion of the printed template for arthroscopy. The endaural approach is marked with an * and the arthroscopic instruments are inserted via the pilot channel. The trocar for cannula or manipulations instruments is inserted via the posterolateral approach (#). The authors recommend a 0° angle for the optic. The central buckle is additional photopolymer/resin and improves the intraoperative handling.

Supplementary Files

This is a list of supplementary files associated with this preprint. Click to download.

- [STROBEchecklistcasecontrol.doc](#)