

Table 3. Contact area reliability for sagittal UO-MRI scans

	Medial Compartment				Lateral Compartment			
	ICC _{3,1} (95%CI)	P-Value	SEM (%)	SEM (mm ²)	ICC _{3,1} (95%CI)	P-Value	SEM (%)	SEM (mm ²)
Inter-Rater	0.95 (0.59-0.99)	0.002	0.39	16.77	0.83 (0.06-0.98)	0.021	0.44	15.48
Test-Retest	0.94 (0.56-0.99)	0.002	0.46	13.33	0.84 (0.10-0.98)	0.017	0.34	11.40
Intra-Rater	0.99 (0.94-1.00)	<0.001	0.21	6.49	0.91 (0.64-0.99)	<0.001	0.31	7.92