

# How to choose the right real-time RT-PCR primer sets for the SARS-CoV-2 genome detection?

**Ahalieyah Anantharajah** (✉ [ahalieyah.anantharajah@uclouvain.be](mailto:ahalieyah.anantharajah@uclouvain.be))

Cliniques universitaires Saint-Luc, Université Catholique de Louvain

**Raphael Helaers**

Human Molecular Genetics, de Duve Institute, Université catholique de Louvain

**Jean-Philippe Defour**

Cliniques universitaires Saint-Luc, Université Catholique de Louvain

**Nathalie Olive**

Cliniques universitaires Saint-Luc, Université Catholique de Louvain

**Florence Kabera**

Cliniques universitaires Saint-Luc, Université Catholique de Louvain

**Luc Croonen**

Cliniques universitaires Saint-Luc, Université Catholique de Louvain

**Françoise Deldime**

Cliniques universitaires Saint-Luc, Université Catholique de Louvain

**Jean-Luc Vaerman**

Cliniques universitaires Saint-Luc, Université Catholique de Louvain

**Cindy Barbée**

Cliniques universitaires Saint-Luc, Université Catholique de Louvain

**Monique Bodéus**

Cliniques universitaires Saint-Luc, Université Catholique de Louvain

**Anais Scohy**

Cliniques universitaires Saint-Luc, Université Catholique de Louvain

**Alexia Verroken**

Cliniques universitaires Saint-Luc, Université Catholique de Louvain

**Hector Rodriguez-Villalobos**

Cliniques universitaires Saint-Luc, Université Catholique de Louvain

**Benoît Kabamba-Mukadi**

Cliniques universitaires Saint-Luc, Université Catholique de Louvain

---

## Research Article

**Keywords:** COVID-19, RT-PCR, SARS-CoV-2

**DOI:** <https://doi.org/10.21203/rs.3.rs-36512/v1>

**License:**  This work is licensed under a Creative Commons Attribution 4.0 International License.

[Read Full License](#)

---

# Abstract

The SARS-CoV-2 pandemic has created an unprecedented need for rapid large-scale diagnostic testing. Currently, several quantitative reverse-transcription polymerase chain reaction (RT-qPCR) assays recommended by the World Health Organization are being used by clinical and public health laboratories and typically target regions of the RNA-dependent RNA polymerase (RdRp), envelope (E) and nucleocapsid (N) coding region. However, it is currently unclear if results from different tests are comparable.

The present study demonstrates substantial differences in SARS-CoV-2 RNA detection sensitivity among the primer/probe sets recommended by the World Health Organization especially for low-level viral loads. The alignment of thousands of European SARS-CoV-2 sequences against the primers/probe highlights single mismatches which might also contribute to false negatives.

An understanding of the limitations depending on the targeted genes and primer/probe sets may influence the selection of molecular detection assays by clinical laboratories.

## Introduction

Efforts to control SARS-CoV-2, the novel coronavirus causing COVID-19 pandemic, depend on accurate and rapid diagnostic testing. The reverse transcription real-time polymerase chain reaction (RT-qPCR) assay has become the gold standard for the diagnosis of SARS-CoV-2 infection. The European Commission recently advised to follow one of the World Health Organization protocols of RT-qPCR assays, published as early as January 2020<sup>1,2</sup>. Among them, the United States Center for Disease Control (US CDC) recommended two nucleocapsid gene targets (*N1* and *N2*)<sup>3</sup> while the German Consiliary Laboratory for Coronaviruses hosted at the Charité in Berlin (Charité/Berlin) recommended first line screening with the envelope (*E*) gene assay followed by a confirmatory assay using the RNA-dependent RNA polymerase (*RdRp*) gene, even before the first COVID-19 cases appeared in Europe<sup>4,5</sup>. At the time of data submission 295 molecular assays are commercially available or in development for the diagnosis of COVID-19 (<https://www.finddx.org/>) and most of them use these recommended gene targets alone or in combination. However, there has been no indication that any one of these sequence regions offer an advantage for clinical diagnostic testing, especially as the number of samples from patients with confirmed COVID-19 has been relatively small in the preliminary evaluations. Recent studies reported that the RT-qPCR assays have limited sensitivity, while chest computed tomography (CT) may reveal pulmonary abnormalities consistent with COVID-19, including ground-glass opacities, multifocal patchy consolidation, and/or interstitial changes with a peripheral distribution, even in patients with negative RT-qPCR results<sup>6-8</sup>. RT-qPCR assays with higher sensitivity might help to reduce the false-negative rate.

In the context of lift confinement restrictions where large scale COVID–19 testing should be needed, this study aimed to clarify the clinical performances of the primer/probe sets designed by US CDC and Charité/Berlin to help guide assay selection by clinical laboratories for SARS-CoV–2 routine detection.

## Results

### Oligonucleotide binding regions alignments

No significant homologies of SARS-CoV–2 sequences with MERS-CoV and seasonal human coronaviruses HKU1, OC43, NL63, 229E were observed suggesting a low risk of potential false positive RT-qPCR results with other circulating human coronaviruses. However, due to the relative paucity of positive control materials at the time these assays were developed, the primers and probe were designed such that they would also cross-react with the SARS-CoV<sup>4,9</sup>. Indeed, probe and primers sequences of *E* gene assay showed high sequence homology with other related betacoronaviruses such as SARS coronavirus and bat SARS-like coronavirus genomes<sup>4</sup>. Although the forward and reverse primer sequences of *RdRp* and *N2* assays showed also high sequence homology with SARS coronavirus, the combination of primers and probe would allow the specific detection of SARS-CoV–2. The assay targeting *N1* gene was found to be specific to SARS-CoV–2.

Very little is known about how rapidly the SARS-CoV–2 genome mutates but its genetic diversity has been reported with a high rate of mutation<sup>10–12</sup>. To assess if these recommended primers and probes covered the circulating strains in Europe, 6750 SARS-CoV–2 sequences were downloaded from GISAID<sup>13</sup>, and compared to the primers and probe binding regions<sup>14</sup> (*Table 1*). Over 99% of analyzed sequences of SARS-CoV–2 showed exact identity to *N1*, *N2* and *E* primers and probe sequences. The remaining sequences of SARS-CoV–2 exhibited 95% of identity corresponding to single base-pair mismatches. It appeared that two types of single mismatches were observed close to the 3' end of the *N1* reverse primer in 31 deposited sequences. A single nucleotide mismatch was found in RdRP\_SARSr-R2 reverse primer with all SARS-CoV–2 analyzed genomes. The primer (CARATGTTAAASACACTATTAGCATA) contained an incorrect degenerate base S at position 12 that binds with G or C while all SARS-CoV–2 analyzed sequences encoded for a T at this position (*Table 1*). This mismatch would not be derived from a new variant but rather due to the initial oligonucleotide design allowing to amplify SARS-CoV, bat-SARS-related CoV and SARS-CoV–2 genomes.

### Clinical performances

We collected nasopharyngeal samples from one hundred patients who presented Chest CT abnormalities consistent with COVID–19 (median age, 63.5 years; 51% female). The average length of stay in COVID–19 care units was 12 days with 15 deaths as of April 30<sup>th</sup> 2020. The nasopharyngeal specimens were obtained upon hospital admission corresponding to an average of 6.6 days after symptoms onset.

Mean Ct values of the sample cohort detected by all RT-qPCR assays were significantly lower in *N1* and *N2* gene assays than in *RdRp*-P2 and *E* gene assay (One way ANOVA, Tukey post-test  $p < 0.001$ ). Compared to CT-Scan, *N1* and *N2* primer/probe sets showed the highest positive rate (73 and 74% respectively) followed by *E* primer/probe set (58%) and then *RdRp* primer/probe sets (44%) (Table 2). The use of Genesig commercial kit with optimized target region and primer/probe sequences in the *RdRp* gene exhibited a slightly increased sensitivity (53%) compare to *RdRp*-P2 assay recommended by Corman et al <sup>4</sup>.

Interestingly, the combination of *N1* and *N2* assays allowed an increase in the sensitivity (84%) compared to *N1* or *N2* alone, including SARS-CoV-2 RNA detection in 5 additional specimens (viral load range: 519 - 1007 copies/mL) that were tested negative by the others assays (Figure 2). Among these patients, we could exclude 3 false positive results as patients had positive SARS-CoV-2 IgG (MAGLUMI assay, Snibe) 7–12 days after molecular testing. Serological control could not be performed for the 2 other patients as they died within a short time spanning.

The performances of the RT-qPCR assays were highly dependent on the viral load. In our study, positive SARS-CoV-2 clinical samples exhibited median Ct values  $> 30$  corresponding to low viral load which made the detection challenging. Indeed, both *RdRp* and *E* assays reliably detected specimens with Ct values  $< 30$ , but did not detect 40–60% of specimens with Ct values  $\geq 30$  (Table 2, Figure 2, Figure S1)..

## Discussion

The survey published on February 11, 2020 reported <sup>5</sup> that the *E*-, *RdRp*- and *N*- gene assays had rapidly been implemented by the European laboratories. Very few studies have been published to date on the relative performance characteristics of these assays recommended by the WHO <sup>3,4,15</sup>. One of the key factors determining detection sensitivity is how efficiently primers and probes bind target genes. Our findings highlight substantial differences in sensitivity for the primer/probe sets when comparing under the same conditions. Indeed, *N1* and *N2* assays stand out in comparison with the *E* and *RdRp* assays for the detection of low-level viral loads. Furthermore, positive *E* and negative *RdRp* results were obtained in 15 cases. We may therefore question the need of confirmatory testing following an initial positive test according to the Charité/Berlin protocol, resulting in turnaround time delay and increased workload. The cross-reactivity of the primers and probes with SARS-CoV should not cause any diagnostic ambiguity as SARS-CoV is no longer detected since the resolution of the epidemic in 2004 <sup>16</sup>. We observed notable mismatches in regions targeted by the primers/probe sets which might affect RT-qPCR assays performance depending on their location and the nature of the substitution <sup>17,18</sup>. Based on our preliminary observations, multiplexing CDC *N1* and *N2* assays within a single PCR mixture could allow a reliable SARS-CoV-2 detection.

At present, in the context of large-scale screening, RT-qPCR testing remains the standard for the diagnosis of COVID-19 despite the false-negative rate. Indeed, normal chest CT scan cannot exclude COVID-19, especially for patients with early symptoms <sup>19</sup> and conversely an abnormal CT scan is not specific for

COVID-19 diagnosis<sup>20</sup>. We believe our results would encourage the laboratory staff to be aware of certain limitations depending on the targeted genes and to evaluate the clinical performances of COVID-19 molecular diagnostic tests across a wide range of viral concentrations before their implementation. Due to the emergence of genetic diversity in the SARS-CoV-2 genome and the extensive transmission of this virus, we also encourage the detection of SARS-CoV-2 by targeting at least two distinct regions and oligonucleotide binding regions should be monitored continuously with the circulating virus strains.

## Methods

### In silico analysis of primer and probe sequences

The amplified regions of SARS-CoV-2 by the recommended primers (E\_Sarbeco\_F1, E\_Sarbeco\_R2, E\_Sarbeco\_P1; RdRP\_SARSr-F2, RdRP\_SARSr-R1, RdRP\_SARSr-P2; 2019-nCoV\_N1-F, 2019-nCoV\_N1-R, 2019-nCoV\_N1-P; 2019-nCoV\_N2-F, 2019-nCoV\_N2-R, 2019-nCoV\_N2-P) were aligned with SARS-Coronavirus, MERS-Coronavirus and seasonal human coronaviruses genome using MUSCLE<sup>21</sup>, and formatted using MSAViewer<sup>22</sup>. Phylogenies have been inferred using MetaPIGA 3.1<sup>23</sup> with “Human CoV 229E” selected as an outgroup. Resulting consensus trees have been formatted using iTOL.

In addition, 6750 SARS-CoV-2 sequences from 33 European countries have been downloaded from Global Initiative on Sharing All Influenza Data (GISAID, as available on 2020 April 24th)<sup>13</sup>, and aligned against *N1*, *N2*, *RdRp* and *E* primers/probe sequences using the Smith-Waterman algorithm<sup>14</sup>.

### Clinical specimens

From April 1 to April 30, 2020, we retrospectively collected one hundred nasopharyngeal swab specimens from patients who were clinically diagnosed with COVID-19 according to the chest CT image and hospitalized in COVID-19 care units of the Cliniques universitaires Saint-Luc, in Brussels, Belgium. An additional 30 “pre-intervention screening” samples from patients who were not suspected of COVID-19 were also included in this study. Nucleic acids from nasopharyngeal swabs were extracted by the Abbott m2000sp following manufacturer’s magnetic microparticle-based protocol (Abbott molecular, IL, USA).

### Real-time RT-PCR assays

The reaction mix (20 µL) consisted of 4x TaqPath™ 1-Step RT-qPCR Master Mix, CG (ThermoFisher), 5 µL of extracted nucleic acid and primers/probes at concentrations recommended by Charité/Berlin and CDC protocols<sup>1</sup>. A multiplexing assay with 2019-nCoV\_N1 and 2019-nCoV\_N2 was also evaluated in parallel. We used human RNase P as control for RNA extraction, and as indicator of the sample quality. The RT-qPCR was performed on LightCycler 480 II (Roche Diagnostics, Mannheim, Germany) with the following amplification parameters: 10 minutes at 55 °C for reverse transcription, 3 minutes at 95 °C for activation

followed by 45 cycles of 15 seconds at 95 °C and 30 seconds at 58 °C <sup>4</sup>. We also tested samples for SARS-CoV-2 using the RdRp genesig® Real-Time PCR COVID-19 CE IVD assay (Primerdesign Ltd, Chandler's Ford, UK) according to the manufacturer's instructions.

## References

- 1WHO. *COVID-19 Technical Guidance: In House Molecular Assays*, <[https://www.who.int/docs/default-source/coronaviruse/whoinhouseassays.pdf?sfvrsn = de3a76aa\\_2&download = true](https://www.who.int/docs/default-source/coronaviruse/whoinhouseassays.pdf?sfvrsn = de3a76aa_2&download = true)> (2020).
- 2European Commission. *Current performance of COVID-19 test methods and devices and proposed performance criteria*, <<https://ec.europa.eu/docsroom/documents/40805>> (2020).
- 3Holshue, M. L. *et al.* First Case of 2019 Novel Coronavirus in the United States. *N Engl J Med* 382, 929–936, doi:10.1056/NEJMoa2001191 (2020).
- 4Corman, V. M. *et al.* Detection of 2019 novel coronavirus (2019-nCoV) by real-time RT-PCR. *Euro Surveill* 25, doi:10.2807/1560–7917.ES.2020.25.3.2000045 (2020).
- 5Reusken, C. *et al.* Laboratory readiness and response for novel coronavirus (2019-nCoV) in expert laboratories in 30 EU/EEA countries, January 2020. *Euro Surveill* 25, doi:10.2807/1560–7917.ES.2020.25.6.2000082 (2020).
- 6Huang, P. *et al.* Use of Chest CT in Combination with Negative RT-PCR Assay for the 2019 Novel Coronavirus but High Clinical Suspicion. *Radiology* 295, 22–23, doi:10.1148/radiol.2020200330 (2020).
- 7Ai, T. *et al.* Correlation of Chest CT and RT-PCR Testing in Coronavirus Disease 2019 (COVID-19) in China: A Report of 1014 Cases. *Radiology*, 200642, doi:10.1148/radiol.2020200642 (2020).
- 8Xie, X. *et al.* Chest CT for Typical 2019-nCoV Pneumonia: Relationship to Negative RT-PCR Testing. *Radiology*, 200343, doi:10.1148/radiol.2020200343 (2020).
- 9Chu, D. K. W. *et al.* Molecular Diagnosis of a Novel Coronavirus (2019-nCoV) Causing an Outbreak of Pneumonia. *Clin Chem* 66, 549–555, doi:10.1093/clinchem/hvaa029 (2020).
- 10van Dorp, L. *et al.* Emergence of genomic diversity and recurrent mutations in SARS-CoV-2. *Infect Genet Evol*, 104351, doi:10.1016/j.meegid.2020.104351 (2020).
- 11Stefanelli, P. *et al.* Whole genome and phylogenetic analysis of two SARS-CoV-2 strains isolated in Italy in January and February 2020: additional clues on multiple introductions and further circulation in Europe. *Euro Surveill* 25, doi:10.2807/1560–7917.ES.2020.25.13.2000305 (2020).
- 12Wang, M. *et al.* International expansion of a novel SARS-CoV-2 mutant. *J Virol*, doi:10.1128/JVI.00567–20 (2020).

- 13Shu, Y. & McCauley, J. GISAID: Global initiative on sharing all influenza data - from vision to reality. *Euro Surveill* 22, doi:10.2807/1560-7917.ES.2017.22.13.30494 (2017).
- 14Smith, T. F. & Waterman, M. S. Identification of common molecular subsequences. *J Mol Biol* 147, 195-197, doi:10.1016/0022-2836(81)90087-5 (1981).
- 15Konrad, R. *et al.* Rapid establishment of laboratory diagnostics for the novel coronavirus SARS-CoV-2 in Bavaria, Germany, February 2020. *Euro Surveill* 25, doi:10.2807/1560-7917.ES.2020.25.9.2000173 (2020).
- 16Wang, M. *et al.* SARS-CoV infection in a restaurant from palm civet. *Emerg Infect Dis* 11, 1860-1865, doi:10.3201/eid1112.041293 (2005).
- 17Klungthong, C. *et al.* The impact of primer and probe-template mismatches on the sensitivity of pandemic influenza A/H1N1/2009 virus detection by real-time RT-PCR. *J Clin Virol* 48, 91-95, doi:10.1016/j.jcv.2010.03.012 (2010).
- 18Stadhouders, R. *et al.* The effect of primer-template mismatches on the detection and quantification of nucleic acids using the 5' nuclease assay. *J Mol Diagn* 12, 109-117, doi:10.2353/jmoldx.2010.090035 (2010).
- 19Yang, W. & Yan, F. Patients with RT-PCR-confirmed COVID-19 and Normal Chest CT. *Radiology* 295, E3, doi:10.1148/radiol.2020200702 (2020).
- 20Hope, M. D., Raptis, C. A. & Henry, T. S. Chest Computed Tomography for Detection of Coronavirus Disease 2019 (COVID-19): Don't Rush the Science. *Ann Intern Med*, doi:10.7326/M20-1382 (2020).
- 21Madeira, F. *et al.* The EMBL-EBI search and sequence analysis tools APIs in 2019. *Nucleic Acids Res* 47, W636-W641, doi:10.1093/nar/gkz268 (2019).
- 22Yachdav, G. *et al.* MSViewer: interactive JavaScript visualization of multiple sequence alignments. *Bioinformatics* 32, 3501-3503, doi:10.1093/bioinformatics/btw474 (2016).
- 23Helaers, R. & Milinkovitch, M. C. MetaPIGA v2.0: maximum likelihood large phylogeny estimation using the metapopulation genetic algorithm and other stochastic heuristics. *BMC Bioinformatics* 11, 379, doi:10.1186/1471-2105-11-379 (2010).

## Declarations

## Acknowledgments

The extraordinary availability of genomic data during the COVID-19 pandemic has been made possible thanks to a tremendous effort by hundreds of researchers depositing SARS-CoV-2 assemblies We acknowledge the authors, originating and submitting laboratories of the sequences from GISAID's EpiFlu



Database on which this research is based (see *Supplementary Excel File*).. We also want to thank for excellent technical assistance the members of the molecular biology laboratory.

## **Authors contributions**

AA, BKM, JPD and RH contributed to the conception and design of the study. AA, FD, NO, FK, LC performed the experiments. RH performed the bioinformatics analysis. JLV and CB contributed to the molecular analysis. AA and BKM analysed all experiments. AA and BKM wrote the manuscript. RH, JLV, MB, AS, AV, and HRV provided critical feedback and contributed to the final version of the manuscript. All authors approved the manuscript.

## **Competing interests**

The authors declare no competing interests

## **Tables**

RT-qPCR assays	Number of analyzed sequences*	Average Position Start sequence	Primer	Primer sequence 5'->3'	Primer size bp	Identity 100%	Identity ~ 95 % Single nucleotide mismatch		
						Number of sequences	Number of sequences	Number of countries	Number of variants
<b>N1</b>	6736	28267	2019-nCoV_N1-F	GACC CCAAA ATCAG CGAAA T	20	6730	6	5	5
			2019-nCoV_N1-P	ACCCC GCATT ACGT TTGG TGGA CC	24	6713	23	8	9
			2019-nCoV_N1-R	TCTG GTTAC TGCC AGTT GAATC TG	24	6672	64	7	12
<b>N2</b>	6724	29144	2019-nCoV_N2-F	TTACA AACAT TGGC CGCA AA	20	6703	21	6	6
			2019-nCoV_N2-P	ACAAT TTGC CCCCA GCGC TTCA G	23	6706	18	5	6
			2019-nCoV_N2-R	GCGC GACAT TCCG AAGA A	18	6715	9	5	6
<b>E</b>	6727	26249	E_Sarbeco_F2	ACAG GTAC GTTAA TAGTT AATAG CGT	26	6716	11	4	8
			E_Sarbeco_P1	ACACT AGCCA TCCTT	26	6710	17	7	5

				ACTG CGCT TCG					
			E_Sarbeco_R1	ATATT GCAG CAGTA CGCA CACA	22	6725	2	2	1
<b>RdRp</b>	6740	15411	RdRP_SARSr-F1	GTGA RATG GTCAT GTGT GGCG G	22	6727	13	8	6
			RdRP_SARSr-P2	CAGG TGGA ACCTC ATCAG GAGA TGC	25	6737	3	1	2
			RdRP_SARSr-R2	CARAT GTAA ASACA CTATT AGCAT A	24	6740	6740	33	1

**Table 1:** Primers alignments on SARS-CoV-2 target sequences

6750 SARS-CoV-2 sequences from 33 European countries (21 Austria, 393 Belgium, 2 Belarus, 7 Croatia, 28 Czech Republic, 337 Denmark, 2444 England, 4 Estonia, 40 Finland, 226 France, 86 Germany, 4 Greece, 3 Hungary, 601 Iceland, 13 Ireland, 58 Italy, 10 Latvia, 1 Lithuania, 204 Luxembourg, 585 Netherland, 1 Northern Ireland, 29 Norway, 15 Poland, 100 Portugal, 105 Russia, 396 Scotland, 4 Slovakia, 4 Slovenia, 190 Spain, 53 Sweden, 76 Switzerland, 30 Turkey, 680 Wales) have been downloaded from GISAID (as available on 2020 April 24th), and aligned against the set of primers/probes. R is G/A and S is G/C.\*Incomplete sequences on target regions were excluded of the analysis

RT-qPCR assays	PCR efficiency (%) <sup>a,b</sup> /linearity (R <sup>2</sup> )	Limit of detection (copies/reaction) <sup>c</sup>	Pre-intervention screening (n=30)	Chest CT-Scan positive				
				All samples (n=100)	Positive samples by <b>all</b> RT-qPCR assays (n=44)		Positive samples by <b>at least one</b> RT-qPCR assay (n=84)	
			Negative agreement % (95% CI)	Positive agreement % (95% CI)	Median Ct (95% CI)	Median Viral Load <sup>a</sup> Log copies/mL (95% CI)	Ct value < 30 (n=19) (95% CI)	Ct value <sup>3</sup> 30 (n=65) (95% CI)
<b>N1</b>	96.81/0.99	5	100 (88.6 – 100)	73 (63.6-80.7)	29.59 (29.8-31.8)	5.07 (4.42-5.89)	100 (83.2 – 100)	83.1 (72.2 – 90.3)
<b>N2</b>	91.62/1.00	5		74 (64.6-81.6)	30.5 (27.92-32.83)	5.06 (4.40-5.79)		84.6 (73.9 – 91.4)
<b>N1+N2</b>	92.85/1.00	5		84 (75.8-89.9)	29.35 (26.82-31.64)	5.11 (4.46-5.83)		100 (94.4 – 100)
<b>RdRp</b>	96.90/1.00	10		44 (34.7-53.8)	31.44 (29.77-33.26)	5.46 (4.92-5.95)		38.5 (27.6 – 50.6)
<b>RdRp Genesig O</b>	80.27/1.00	10		53 (43.3-62.5)	31.67 (28.87-33.41)	4.47 (4.02-5.18)		52.3 (40.4 – 64.0)
<b>E</b>	95.68/1.00	5		58 (48.2-67.2)	31.06 (28.25-33.76)	4.98 (4.19-5.79)		60.0 (47.8 – 71.0)

**Table 2:** Comparison of the six RT-qPCR assays performances

<sup>a</sup> PCR efficiency, linearity and virus copies were determined using a 10-fold dilution standard curve of positive control plasmids.


N1 :  $y = -3.4007x + 36.624$ ; N2 :  $y = -3.5407x + 37.783$ ; N1+N2 :  $y = -3.506x + 36.74$  ;

RdRp :  $y = -3.3985x + 39.798$  ; RdRp Genesig $\hat{O}$  :  $y = -3.9076x + 37.404$  ; E :  $y = -3.43x + 37.84$

<sup>b</sup> PCR Efficiency E =  $100 * (10^{-1/\text{slope}} - 1)$ .

<sup>c</sup> The Limit of detection was determined as the lowest concentration where  $\geq 90\%$  (9/10) of the replicates were positive

## Figures



**Figure 1**

Alignments of N1 (A), N2 (B), E (C), RdRp (D) regions of SARS-CoV-2 with SARS-Coronavirus, MERS-Coronavirus and seasonal human coronaviruses genomes. The arrows indicate the regions targeted by the set of primers/probes. Human SARS-CoV-2: hCoV-19/Belgium/CJM-0323175/2020|EPI\_ISL\_420432; Human SARS-CoV: SARS coronavirus NC\_004718.3; Bat SL-CoV ZXC2: Bat SARS-like coronavirus isolate MG772934.1; Bat SARS-related CoV BM48-3: BGR/2008 GU190215.1 Human MERS-CoV: Middle East respiratory syndrome coronavirus NC\_019843.3; Human CoV HKU1: Human coronavirus HKU1 NC\_006577.2; Human CoV OC43 : Human coronavirus OC43 strain ATCC VR-759 NC\_006213.1; Human CoV NL63: Human Coronavirus NL63 NC\_005831.2; Human CoV 229E: Human coronavirus 229E NC\_002645.1

### All positive chest CT scan & positive RT- qPCR samples




Figure 2

Comparison of the viral load detected by the six RT-qPCR assays among the positive nasopharyngeal swabs ( $n = 84$ ). The viral load is expressed in log copies/mL and each clinical sample is represented by a circle. The white circles represent clinical samples detected by all RT-qPCR assays while colored circles represent samples not detected by the six assays. Bars represent the median and 95% Confidence Interval

## Supplementary Files

This is a list of supplementary files associated with this preprint. Click to download.

- [gisaidhcov19acknowledgementtable2020051916.xls](#)
- [FigureS1.pdf](#)